



## Weekly MD Report – 13 to 17 October 2014 – Gian Piero Di Giovanni

On Monday and Tuesday I managed to get some time to use both an ISOGPS\_Clone beam for measurements in LT/LTB/BI and an EASTA\_Clone beam to be sent from ring 3 to the PS.

Concerning the injection transfer line, I verified the linearity of the BPMs to make sure the disagreement I have been observing so far were not a simple artifact of the hardware. I could find a linear response for all the BPMs in LT/LTB/BI.



By analyzing the BT-BTP line, I could confirm the problem observed last week: BTP.BPM10 horizontally seems to be insensitive to any kick. BI experts are informed and will look at it on Friday morning.



I also got measurements for the kick response in the BTM line, but I did not have the time to analyze them yet.